

# V&V Summary Report

## L2 ASCDS Version : 7.6.8.1

Observation 3891 - L2 Version 001  
Chandra X-Ray Center

L2 Processing Date : Aug 10 2006

See axaff03891N001\_VV001\_vvref2.pdf for the full report

V&V Scientist	Jen Lauer
V&V Date (YYYY-MM-DD)	2006.08.10
V&V Edition	1
V&V Disposition and Status	OK
V&V Charge Time	19.291

## Comments

The ACIS focal plane temperature was above calibration limits for the first 2 ksec of the observation.

The ACIS spectral response calibration is less accurate at these warmer temperatures than it is at -119.7 C. Users whose science objectives depend on the most accurate spectral response (ie: fitting line-rich spectra) may notice an effect. Users whose science objectives do not depend on the most accurate spectral response should not notice an effect.

seq_num	500370
obs_id	3891
title	IS AX J1845-0258 AN ANOMALOUS X-RAY PULSAR?
observer	Professor Victoria Kaspi
object	AX J1845-0258
dtcycle	0
cycle	P
ra_targ	281.2275
dec_targ	-2.948056
ra_nom	281.21885027151
dec_nom	-2.9437067647324
roll_nom	163.23861308496
revision	2
ontime	19291.877187401
livetime	17496.714300201
ontime7	19291.877187401
l2events	8253

