## V&V Summary Report L2 ASCDS Version: 8.1.2

Observation 1061 - L2 Version 3 Chandra X-Ray Center

L2 Processing Date: Dec 14 2009

See axaff01061N001\_VV001\_vvref2.pdf for the full report

V&V Scientist	Joy Nichols
V&V Date (YYYY-MM-DD)	2010.05.19
V&V Edition	1
V&V Disposition and Status	OK
V&V Charge Time	1.76

## Comments

ACIS-S Crude Optical Axis Determination Yoff = -3.0, Zoff =-2.8. Target

very off-axis.

===

Guide star in slot 4 was not acquired. The guide star in slot 3 was removed from the aspect solution due to poor data quality.

===

The focal plane temperature is approximately -100 C during this observation. This reprocessing of the data applies no CTI correction because none is available for this temperature at present.

The ACIS CTI correction has not been calibrated at this temperature, because it was early in the mission, and ACIS had not yet been lowered to the standard -119.7 C. Both front and back illuminated chips are affected. However a  $T_GAIN$  correction has been applied to the BI chips (ACIS-5 and ACIS-7) data included here.

The ACIS spectral response calibration is less accurate at these warmer temperatures than it is at -119.7 C. Users whose science objectives depend on the most accurate spectral response (ie: fitting line-rich

depend on the most accurate spectral response should not notice an effect.

1	
780011	Sequence number
1061	Observation id
<b>&amp;</b> #160	Proposal title
Dr. CXC Calibration	Principal investigator
PKS0637-752	Source name
0	<b>&amp;</b> #160
P	events from which exps? Prim/Second/Both
98.94	Observer's specified target RA
-75.27	Observer's specified target Dec
99.201266156527	Nominal RA
-75.267203296951	Nominal Dec
135.12468443304	Nominal Roll
3	Processing version of data
1758.3858210817	Sum of GTIs [s]
1736.1200810423	Livetime [s]
1758.5089410841	Sum of GTIs [s]
1758.4679010808	Sum of GTIs [s]
1758.54998108	Sum of GTIs [s]
1758.3858210817	Sum of GTIs [s]
1758.5910210833	Sum of GTIs [s]
1758.426861085	Sum of GTIs [s]
45293	Number of level 2 events
	1061  Dr. CXC Calibration PKS0637-752 0 P 98.94 -75.27 99.201266156527 -75.267203296951 135.12468443304 3 1758.3858210817 1736.1200810423 1758.5089410841 1758.4679010808 1758.54998108 1758.3858210817 1758.5910210833 1758.426861085

